目 录

1	目录
2-5	质检报告（英文）
6-9	质检报告（中文）
10	营业执照
11	CE认证
12	商标注册证
13	fda认证

CMA 180011112242

（2018）国认监认字（244）号

ilac-MRA

CNAS 中国认可 国际互认 检测 TESTING CNAS L6780

检 验 报 告

TEST REPORT

Scan is following us

Scan checks for authenticity

W202007502E

浙江省轻工业品质量检验研究院

Zhejiang Light Industrial Products Inspection and Research Institute

国家纺织服装产品质量监督检验中心（浙江）

National Textiles and Garment Quality Supervision Inspection Center(Zhejiang)

Zhejiang Light Industrial Products Inspection and Research Institute
National Textiles and Garment Quality Supervision Inspection Center(Zhejiang)

Test report

Number: W202007502E

Name of Customer	Cangnan YiLong clothing Co.,Ltd	Address	Building 5, small business pioneer park, Qianku Town, Cangnan County, Wenzhou City, Zhejiang Province
Manufacturer	——	Address	——
Sample information	Name of sample: KN95 Mask Characteristics of sample: white Trademark of sample: —— Specification/model: —— Level: KN95 Category of safety specification: —— Art. No.: ——		
Above-mentioned information by Customer-supplied			
The sent way of sample	Courier	Sample quantity	30 pieces
Receiving Date of Sample	2020/04/18	Test Category	Entrusted inspection
Date of Testing	2020-04-18- 2020-04-18		
Rating Requirements	GB 2626-2006		

Test Summary:

See the attached page for the results.

Test Seal

Date of Approval: 2020-04-22

Remarks	The sample is not pretreated and the result is for reference.

Approved by: 曹丽勤

Test report

Number:W202007502E

ITEM	STANDARD (KN95)	RESULT	RATING
1. FILTER EFFICIENCY(SALT MEDIUM) (GB 2626-2006)(%)			
-	≥95.0	97.7	PASS
2. RESPIRATORY RESISTANCE (GB 2626-2006)(Pa)			
-INSPIRATORY RESISTANCE	≤350	101	PASS
-EXPIRATORY RESISTANCE	≤250	83	PASS
3.APPEARANCE (GB 2626-2006)			
-	Meet the requirements of GB 2626-2006 standard	The appearance quality of 2 masks conforms to the standard	PASS
4. VISUAL FIELD BELOW MASK (GB 2626-2006、GB 2890-2009)(°)			
-	≥60	78	PASS
5. HEADBAND (GB 2626-2006)			
	≥10N,hold on10s	No breakage and slippage were observed	PASS

Test report

Picture(s) of sample





—End of report—

CMA 180011112242　CAL （2018）国认监认字（244）号　ilac-MRA　CNAS 中国认可 国际互认 检测 TESTING CNAS L6780

检 验 报 告

TEST REPORT

扫一扫关注我们　扫一扫查询真伪

报 告 编 号 REPORT NO.	国纺委字第 W202007502 号
产 品 名 称 NAME OF SAMPLE	KN95
委 托 单 位 CUSTOMER	苍南亿龙服饰有限公司
检 验 类 别 TEST CATEGORY	委托检验

浙江省轻工业品质量检验研究院

（浙江省纺织测试研究院）

Zhejiang Light Industrial Products Inspection and Research Institute

国家纺织服装产品质量监督检验中心（浙江）

National Textiles and Garment Quality Supervision Inspection Center(Zhejiang)

浙江省轻工业品质量检验研究院
国家纺织服装产品质量监督检验中心（浙江）

检验报告

国纺委字第 W202007502 号　　第 1 页 共 3 页

委托单位名称 Name of Customer	苍南亿龙服饰有限公司	地 址 Address	/
生产单位 Manufacturer		地 址 Address	——
样品信息 Sample information	样品名称 Name of sample：KN95 样品特性 Characteristics：白色 商标 Trademark：—— 规格/号型 Specification/model：—— 等 级 Level：KN95 安全技术类别 Category of safety specification：—— 样品款号/货号 Art. No.：——		
以上为客供信息（Above-mentioned information by Customer-supplied）			
来样方式 The sent way of sample	自送	样品数量 Sample quantity	30 只
送检日期 Receiving Date of Sample	2020-04-18	检测类别 Test Category	委托检验
判定依据 RatingRequirements	GB 2626-2006		

检测结论/Test Summary：

实测结果详见附页。

（检验报告专用章/Test Seal）

批准日期/ Date of Approval：2020-04-19

备 注 Remarks	样品未经预处理，数据供参考。

签 发：
Approved by

检验报告

国纺委字第 W202007502 号

序号	检测项目		检测方法	单位	标准要求（KN95）	实测值	单项评价	结果备注
1	过滤效率（盐性介质）		GB 2626-2006 6.3	%	≥95.0	97.7	符合	——
2	呼吸阻力	吸气阻力	GB 2626-2006 6.5 6.6	Pa	≤350	101	符合	——
		呼气阻力		Pa	≤250	83		
3	头带		GB 2626-2006 6.11	——	≥10N，保持 10s	未出现断裂和滑脱现象	符合	——
4	外观检查		GB 2626-2006 6.1 6.2	——	符合 GB2626-2006 标准要求	2 个口罩的外观质量符合标准	符合	——
5	口罩下方视野		GB 2890-2009 6.8	°	≥60	78	符合	——

检验报告

样品图片





—以下空白—

统一社会信用代码

91330327079723 93X5

营业执照

扫描二维码登录"国家企业信用信息公示系统"了解更多登记、备案许可、监管信息

名　　称　苍南亿龙服饰有限公司

类　　型　有限责任公司（自然人投资或控股）

法定代表人　黄苏莲

经营范围　一般项目：服装制造，针纺织品销售，日用口罩（非医用）生产，日用口罩（非医用）销售（除依法须经批准的项目外，凭营业执照依法自主开展经营活动）。

注册资本　壹佰万元整

成立日期　2013年09月26日

营业期限　2013年09月26日 至 长期

住　　所　浙江省温州市苍南县钱库镇箱包小微型创业园5幢

登记机关　苍南县市场监督管理局

2020 年 03 月 05 日

国家企业信用信息公示系统网址：http://www.gsxt.gov.cn/

国家市场监督管理总局监制

Certificate – Сертификат – 證明書 – Certificat – 증명서 – شهادة

Certificate of Compliance

No. 4T200407F.CYC0O77

Certificate's Holder: Cangnan YiLong clothing Co.,Ltd
Building 5, small business pioneer park, Qianku Town, Cangnan County, Wenzhou City, Zhejiang Province

Certification ECM Mark: ENTE CERTIFICAZIONE MACCHINE – European Conformity – SAFETY COMPLIANCE MARK – Type Approved

Product: Protective mask

Model(s): Disposable protective mask
KN95

Verification to: Standard:
EN 149:2001+A1:2009

related to CE Directive(s):
R 2016/425 (Personal Protective Equipment)

Remark: This document has been issued on a voluntary basis and upon request of the manufacturer. It is our opinion that the technical documentation received from the manufacturer is satisfactory for the requirements of the ECM Certification Mark. The conformity mark above can be affixed on the products accordingly to the ECM regulation about its release and its use.

Additional information and clarification about the Marking:

CE

The manufacturer is responsible for the CE Marking process, and if necessary, must refer to a Notified Body. This document has been issued on the basis of the regulation on ECM Voluntary Mark for the certification of products. RG01_ECM rev.3 available at: www.entecerma.it

Issuance date: 07 April 2020

Expiry date: 06 April 2025

Reviewer
Technical expert
Amanda Payne

Approver
ECM Service Director
Luca Bedonni

Ente Certificazione Macchine Srl
Via Ca' Bella, 243 – Loc. Castello di Serravalle – 40053 Valsamoggia (BO) - ITALY
☎ +39 051 6705141 🖷 +39 051 6705156 ✉ info@entecerma.it 🌐 www.entecerma.it

TMZC30194168D01T190313

第 30194168 号

商标注册证

COCOTAN

核定使用商品/服务项目（国际分类：10）

第10类：振动按摩器；避孕套；健美按摩设备；畸齿矫正仪器；矫形用物品；医用熏蒸设备；婴儿用安抚奶嘴；电热敷布（外科）；腹部护垫；口罩（截止）

注 册 人 深圳市千里度科技有限公司

注册人地址 广东省深圳市龙岗区坂田街道中兴路143号鸿和达大厦A座301

注册日期 2019年02月07日 有效期至 2029年02月06日

局 长

发证机关 国家知识产权局



Fiscal Year 2020
CERTIFICATION OF REGISTRATION

This certifies that:

Cangnan YiLong clothing Co.,Ltd.
Building 5, small business pioneer park, Qianku Town, Cangnan County， Wenzhou City, Zhejiang Province

has completed the FDA Establishment Registration and Device Listing.

Owner/Operator Number:10071319

Listing No.	Code	Device Name	Proprietary Names
D395123	QKR	Face mask (except N95 respirator) for general public/healthcare personnel per IIE guidance	Disposable Face Mask 175mm*95mm
D395124	MSH	Respirator, surgical	KN95 Mask 160mm*105

Initial Registration Date: April 22, 2020
Expiration Date: December 31, 2020